

Certificate of Analysis

Certification Date:	15 December 2017
Product:	Broad Spectrum Microtab II - Bulk
Part Number:	954
Lot Number:	21117
Date of Manufacture:	NOV 2017
Expiration Date:	NOV 2021

Product Data

<u>Test Method</u>	<u>Analytical Result</u>	<u>Specification</u>
Product Description	Pass	A small, round, pink/mauve, biconvex tablet
Average Weight	18.24 mg	17.00 mg - 19.00 mg
Diameter	3.00 mm	2.94 mm - 3.06 mm
Solution Time (pasteurized milk)	14 minutes	Not more than 30 minutes
Solution Appearance	Pass	No specks visible
Bronopol Content	7.06 mg/tablet	7.00 mg - 9.00 mg
Natamycin Content	0.28 mg/tablet	0.25 mg - 0.35 mg

The Quality Management Systems of Advanced Instruments, Inc. is registered as conforming to the requirements of ISO 9001 and ISO 13485.

Analyst
 Printed Name: MEREDITH PESTA
 Signature: Meredith Pesta

Approver
 Printed Name: Mark Hunter
 Signature: M. Hunter